



📍 Overland Park, KS
📞 620-805-1216
✉ tywon@h2hubb.com
🌐 www.h2hubb.com

H₂ HUBB Official Test Report

Product:

Name: AquaCure AC50

Company: Eagle Research

Type: Oxyhydrogen Inhalation Device (66%: H₂ / 33%: O₂)

- Electrolytic Cell Type
 - No membrane

Model: AC50-B

Tester: Tywon Hubbard (TH)

Testing start date: 6/22/19

Completion date: 8/13/19

PERFORMANCE:

H₂ mL/min Confirmation Test: VH600

- **METHODOLOGY:**
- Distilled water (used for testing): 6.0 pH
- Electrolysis Electrolyte: Sodium Hydroxide: NaOH (crystal form)
- 150 g/L (4M)
- Water Temperature: 61~66F/ 16.1~18.8C
- Reservoir Size: 1.2 Liters (1200 ml)
- Preferred Reservoir Vol Level: less than 80% full (>960 ml or 30.4 oz)
- H₂ output: 100% Capacity: > 600 mL/min or 49.5 mg/min (@ SATP)
- Test Location: 277 meters (909 ft elevation)
- H₂ Flow Test: mL/min, normal timing for breathing session
 - Test methodology: Gas Displacement
 - All measurements converted to SATP
- Tests performed at 40% output capacity: Recommended inhalation level
- Tests performed at 100% output capacity: Maximum output

H₂ Flow Test at SATP:

- **Device H₂ mL/min (mg/min) avg:**
 - 40% Capacity: ≈ 394 mL/min (32.5 mg/min)
 - 100% Capacity: ≈ 640.55 mL/min (52.81 mg/min)
- **Claimed Mfgr's H₂ mL/min (mg/min) confirmed:** Yes

Summary Report Only.

Not Full Test Report.

Other testing and technical sections are not included out of respect and professional courtesy of the RPC.

H₂ Hubb LLC disclaimer: All tests conducted and test results produced by H₂ Hubb LLC have been done according to industry-accepted practices and standards. Nevertheless, these results may not necessarily reflect test results performed by manufacturers, suppliers or third-party labs. Our test results are independent of all other parties, and testing by other parties may produce different results. We understand that many variables are involved in testing, some of which are extremely difficult to control. These reports are not meant or intended for any other purpose but to uphold H₂ Hubb LLC business practices and to validate the reasons for our recommendations.

Approved by: Tywon Hubbard

A handwritten signature in black ink, appearing to read 'Tywon Hubbard' in a cursive style.

Tywon Hubbard,
CEO, H₂ HUBB LLC.

Tywon@H2HUBB.com