



Overland Park KS
Tywon@H₂HUBB.com
www.H₂HUBB.com

Date: 8/19/2025

H₂HUBB Official Test Report

Evaluation Introduction

This report provides a comprehensive analysis of the Gaura Silmare Pulse 120 Hydrogen Inhalation Device from VHLife, a company based in Taiwan. H₂HUBB classifies this device as a portable, household-grade, low-to-mid-flow system. The unit uses a PEM/SPE electrolytic cell to produce pure hydrogen gas and requires only deionized or distilled water. It supports continuous delivery and a breath-synchronized pulse mode with auto start/stop; sessions can run up to 2 hours, and in Auto mode the device continues producing hydrogen while the user is inhaling. We quantified hydrogen output (mL/min) to verify performance and assessed safety features and operating controls to confirm appropriate protections for safe, reliable use. This report determines whether the Gaura Silmare Pulse 120 meets H₂HUBB's product performance standards for approval and recommendation. For more information about our performance standards for hydrogen inhalation systems, please visit [H₂HUBB](https://www.h2hubb.com).

H₂ Products

- Company: VHLife
- Product Name: Gaura Silmare Pulse 120
- Type: Pure H₂ Inhalation Device (99.995%/4N)
 - PEM/SPE
 - low-to-mid Flow Rate
 - Pulse mode
- Mfg rated H₂ Output: 120 mL/min
- URL Link: <https://www.vhlife.com/zh/>

Method and Procedure

- Distilled Water (used for testing): 6.0 pH
- Water Temperature: 65~70F/ 18~21C
- Reservoir Vol Size: 0.350 L/350 mL (11.83 oz)
- H₂ Output: 120 mL/min or 9.89 mg/min (@ SATP)
- Pulse mode equivalent: 360 mL/min or 29.68 mg/min (@ SATP)
- Test Location: 277 meters (909 ft elevation)
- H₂ Flow Test: mL/min, normal timing for a breathing session (1 hr)
- Test methodology:
 - [Alicat H₂ Mass Flow Meter](#)
 - [Gas Displacement](#)
- All measurements converted to SATP where applicable

Test Results

To measure hydrogen gas flow rate in continuous and pulse modes, the device was assembled per manufacturer guidance and filled with distilled water to the recommended level. The operating mode (normal or pulse) was selected via the mode button, and each condition was run for a 1-hour session with a 10-minute warm-up preceding measurement. Product gas was routed through a dryer, then a combined humidity/temperature sensor, and finally into an Alicat mass flow meter for quantification of molecular hydrogen output. For pulse mode, a calibrated gas-displacement apparatus was additionally employed to determine per-pulse volume and corroborate average pulse-mode flow. Before recording, the system was allowed 5–10 minutes to reach steady state to minimize mixing with ambient air. Minor calculations and corrections were applied to account for residual moisture, inline resistance from the dryer, and volumetric/thermal effects during pulse capture; reported values reflect these adjustments. A minimum of three tests were conducted per condition, and the reported flow-rate values are the average of those runs.

H₂ Flow Rate Test Results at SATP:

- Device H₂ output, avg: $\cong 150 \text{ mL/min}$ (2.5 mL/s) $\approx 12.38 \text{ mg/min}$
- Device O₂ output (anode, vented): $\cong 75 \text{ mL/min} \approx 98 \text{ mg/min}$
- Total electrolysis production (H₂ + O₂): $\cong 225 \text{ mL/min}$

- Pulse Mode (no breathing / reservoir behavior)
- Pulse interval (no inhalation): $\cong 6\text{--}6.5 \text{ s}$
- H₂ available per pulse (calculated): $\cong 15\text{--}16.25 \text{ mL}$ ($2.5 \text{ mL/s} \times 6\text{--}6.5 \text{ s}$)
- H₂ released per pulse (measured): $\cong 13\text{--}15 \text{ mL}$

- Pulse Mode (while breathing / GMMS breath-sync)
- Breath interval: \cong every 3–4 s
- H₂ per inhalation (calculated): $\cong 7.5\text{--}10 \text{ mL}$ ($2.5 \text{ mL/s} \times 3\text{--}4 \text{ s}$)
- H₂ per inhalation (observed): $\cong 7.5\text{--}10 \text{ mL}$
- Breath-trigger reliability: $\cong 90\text{--}100\%$ of inhalations

- **Claimed Mfg'r's H₂ mL/min (mg/min) confirmed: Yes**

H₂HUBB Hydrogen Flow Rate Assessment

- H₂HUBB's hydrogen gas flow-rate testing confirms that the Gaura Silmare Pulse 120 meets—and modestly exceeds—the manufacturer's stated output. Under SATP conditions, the device produced 150 mL/min of hydrogen in continuous operation, with consistent results across repeated runs, indicating stable slight over-delivery relative to the 120 mL/min claim. In pulse mode, breath-synchronized delivery concentrates production into inhalation, producing approximately 7.5–10 mL H₂ per breath for the vast majority of inhalations, which increases the effective inspired fraction of hydrogen by roughly three- to four-fold compared with continuous flow. Using a nasal cannula and a minute ventilation (VE) of 5–6 L/min, this corresponds to an average inspired hydrogen concentration (FiH₂) of ~1.0% (5 L/min) or ~0.83% (6 L/min) in continuous mode versus ~3.0–4.0% (5 L/min) or ~2.5–3.3% (6 L/min) in pulse mode. These results exceed **H₂HUBB's minimum performance standards** and qualify the Gaura Silmare Pulse 120 as a **Level 3 hydrogen inhalation device** within our performance ranking system. The first ever H₂HUBB approved portable H₂ inhalation device to receive this rating.

INTERNAL BREAKDOWN AND PERFORMANCE:

Manufacturer's Rated Electrical Values:

- Type of device/electrolytic cell
 - Pure H₂: PEM/SPE membrane
- AC input: AC100–240 V, 50–60 Hz
- Maximum power: 60 W
- Rated H₂ flow: 120 mL/min (spec)

GMMS pulse system (manufacturer-listed)

- GMMS: respiration detection with auto start/stop
- Inhalation-triggered pulse delivery (releases H₂ only during inhale)
- Claimed intake-efficiency multiplier: ~2.5–3.5× vs normal/continuous mode
- Hydrogen “pulse efficiency” (manufacturer claim): $\approx 360 \text{ mL/min}$ (continuous-equivalent)

Product Assessment

Functionality:

- **Power input/Power cord:**
 - Rear-mounted AC inlet; supplies power to the device (AC100–240 V, 50–60 Hz; grounded outlet recommended).
- **LED Digital Display and Control Panel:**
 - Start/Stop/Timer button and Mode/Sleep button.
 - Shows timer status and alarm indicators (water level/quality, temperature, pressure).
- **Output Flow:**
 - Pure H₂ via PEM/SPE cell.
 - Delivery modes: Normal (continuous) or Pulse (GMMS breath-sync).
 - Rated H₂ flow: 120 mL/min (H₂HUBB measured ≈150 mL/min at SATP).
- **Power Start/Stop Button:**
 - Starts electrolysis; long-press (~3 s) to stop and enter standby.
- **Timer:**
 - Auto (pulse mode with auto start/stop on detected breathing), 1-hour, or 2-hour sessions.
 - Last setting retained in standby.
- **Pulse Mode (GMMS):**
 - Detects inhalation and releases stored H₂ only during inhale to reduce waste.
 - Auto start/stop when cannula is put on/removed.
- **Sleep Mode:**
 - Press Mode for ~3 s to dim the display and touch panel; press again to wake.
- **Reservoir (~0.35 L / 11.83 oz):**
 - Fill to marked level with purified/distilled water; low TDS required (manual: ≤ ~5–10 µS/cm).
 - Front water-level window.
- **H₂ Outlet (1x):**
 - For single-user inhalation; includes silicone inhalation tube with dryer.
- **O₂ Vent (1x):**
 - Vents oxygen generated during electrolysis; keep unobstructed (remove plug when operating).
- **Drain Port:**
 - Bottom drainage outlet/plug to empty the tank before transport or maintenance.
- **Safety/Alarms:**
 - Low water level, poor water quality (high TDS), H₂ route blockage/pressure, high electrolyzer temperature.
 - Device auto-stops on alarms or when tubing is removed (auto mode).

Product Safety

Safety Components:

- The system has 9 fundamental safety mechanisms for ensuring the device's safety.
 - Low-water protection
 - Protects cells from excessive heat
 - Distilled water reservoir
 - Maintains purity; limits heat stress.
 - Timer/idle protection
 - 1 h / 2 h session limits.
 - Auto start/stop on breathing (GMMS)
 - Starts on inhale; stops if cannula off (~15 s).
 - Water-quality (TDS) alarm with auto stop
 - Auto-stop when water quality is out of spec.
 - Hydrogen line blockage/over-pressure alarm
 - Warns/halts when the outlet is obstructed.
 - H₂ cell temperature alarm
 - Shuts down on high cell temperature.
 - Internal Fans
 - Vent gas; aid cooling.
 - Heat Vents
 - Dissipate internal heat.

The system theoretically should only be combustible at the tip of the nasal cannula as the system produces >99% pure hydrogen gas. As with all inhalation devices that produce pure hydrogen gas, care should be taken to avoid exposing the cannula tip to any source of ignition (such as an open flame or a spark) which could result in the combustion of the gas.

Overall Opinion

The Gaura Silmare Pulse 120 Hydrogen Inhalation Device has been confirmed through our testing as a well-engineered system for hydrogen inhalation. The manufacturer rates the device to deliver 120 mL/min of 99.995% pure hydrogen gas under standard operating conditions. Our testing verified these claims, demonstrating that the device consistently achieves the specified hydrogen output, with measured results aligning with the rated flow rate. Additionally, our evaluation confirmed the functionality of the device's Pulse Mode, which enhances per-breath hydrogen delivery and effectively increases inhaled hydrogen concentration in alignment with the manufacturer's intended design.

Hydrogen gas output flow rates are a critical performance parameter for inhalation devices. At H₂HUBB, the minimum standard for hydrogen generators or inhalation units (whether pure hydrogen, oxyhydrogen, or H₂ mixed with air) is 120 mL/min of H₂. This rate corresponds to approximately 0.7-1.0% H₂ at a resting breathing rate of 4-6 L/min when using a nasal cannula for an average adult. Scientific studies on molecular hydrogen inhalation therapy generally utilize concentrations between 0.5% and 4% or more at resting breathing rates, a range that has been shown to provide therapeutic benefits. Given these findings, H₂HUBB establishes 120 mL/min of H₂ as the baseline requirement for hydrogen inhalation devices to ensure effectiveness. The Gaura Silmare Pulse 120 H₂ inhalation device significantly exceeds this minimum standard, delivering performance within the therapeutic range.

The Gaura Silmare Pulse 120 uses a PEM/SPE electrolytic stack configured as two cells in series to produce high-purity hydrogen (>99.99%). In our testing at SATP, the device delivered 150 mL/min of H₂ (~12.37 mg/min) in continuous operation. Using our electron-count method (1 A ≈ 7.6 mL/min H₂ per cell), a two-cell series stack producing 150 mL/min corresponds to an operating current of ≈10.5 A. Within the 60 W nameplate, the estimated DC operating point at the stack is ≈4.0–5.0 V (≈2.0–2.5 V per cell), or ≈40–50 W. At 10.5 A, the theoretical 100%-efficiency output is ~160 mL/min; our measured 150 mL/min indicates ~94% cell efficiency and a ~25% over-delivery relative to the manufacturer's 120 mL/min specification. This suggests the manufacturer has taken a conservative approach in its specifications, opting to under-promise and over-deliver—an approach that deserves commendation. These values support classifying the Silmare Pulse 120 as a portable, household-grade, low-to-mid-flow hydrogen inhalation system. **According to our flow-rate testing, we will list this device on our website as a Level 3 hydrogen inhalation system, specifically due to its GMMS breath-synchronized pulse mode; details follow in the next section.** You can view the meaning of this ranking [here](#).

GMMS Pulse Mode (Gaura Silmare Pulse 120)

The GMMS pulse mode releases hydrogen only during inhalation to reduce waste between breaths. After selecting Pulse with the Mode button, the system detects each inhale (negative pressure at the cannula) and delivers a small stored dose of H₂. Based on our testing, with a measured production of 150 mL/min (≈2.5 mL/s), typical per-breath delivery is ~7.5–10 mL at ~3–4 seconds between breaths, and up to ~12–13 mL at ~5 seconds between breaths (~12 bpm). When no one is inhaling, we observed a reservoir purge of ~13–15 mL every ~6–6.5 seconds, implying a reservoir capacity of ~15–16 mL. Based on our testing, breath-trigger reliability was ~90–100% of inhalations.

Minute ventilation (VE) is the total air you move each minute (generally 4–8 L/min at rest). FiH₂ (inhaled hydrogen percentage) is the amount of hydrogen that actually reaches the lungs during inhalation each minute divided by VE. Based on our testing, in continuous mode about 50 mL of H₂ reaches the lungs during inhalation each minute (we generally inhale ~20 seconds out of 60), so $FiH_2 \approx 50 \div VE$ (e.g., $50/5000 \text{ mL} = 1\%$ inhaled H₂). In GMMS pulse mode, H₂ is delivered only in the inhalation window, allowing about 150 mL of H₂ per minute to reach the lungs during inhalation, so $FiH_2 \approx 150 \div VE$ (e.g., $150/5000 \text{ mL} = 3\%$ inhaled H₂). Functionally, this concentrates the same minute production (150 mL/min of H₂) into the ~33% inhalation window (~20 seconds of inhalation per minute at rest), which raises the inspired hydrogen concentration about 3–4 times versus continuous flow at the same VE. Using $VE = 5\text{--}6 \text{ L/min}$, this corresponds—based on our testing—to ~1.0% (5 L/min) or ~0.83% (6 L/min) in continuous mode versus ~3.0–4.0% (5 L/min) or ~2.5–3.3% (6 L/min) in pulse mode. These findings are consistent with our per-pulse gas-displacement measurements and continuous-flow verification.

Below you will find a table for FiH₂ versus minute ventilation (VE) up to 8 L/min at a resting breathing pattern of ~12 breaths per minute (~5 seconds per breath; ~3–4 seconds between inhalations depending on I:E).

Inhaled H₂ (%FiH₂) vs Minute Ventilation (VE): Continuous vs GMMS Pulse

VE (mL/min)	Continuous FiH ₂	Pulse FiH ₂ (7.5–10 mL per breath @ ~20 bpm)
4,000	1.25%	3.75–5.00%
5,000	1.00%	3.00–4.00%
6,000	0.83%	2.50–3.33%
7,000	0.71%	2.14–2.86%
8,000	0.71%	1.88–2.50%

The manufacturer’s hydrogen output claim for the Gaura Silmare Pulse 120 was validated by our testing under SATP; we measured 150 mL/min, modestly exceeding the 120 mL/min specification. Based on our testing, GMMS pulse mode concentrated delivery into inhalation and produced a 3–4× increase in inspired H₂ fraction at typical resting VE (5–6 L/min), consistent with the device’s operating logic. We identified no safety concerns during evaluation; the system’s protections (low water, water quality, pressure, temperature, and auto start/stop on breathing) functioned as intended. Overall, the device meets H₂HUBB’s minimum performance standards and is suitable for in-home hydrogen inhalation. We are confident in recommending the Gaura Silmare Pulse 120 to the public.

H₂ Hubb LLC disclaimer: All tests conducted and test results produced by H₂ Hubb LLC have been done according to industry-accepted practices and standards. Nevertheless, these results may not necessarily reflect test results performed by manufacturers, suppliers or third-party labs. Our test results are independent of all other parties, and testing by other parties may produce different results. We understand that many variables are involved in testing, some of which are extremely difficult to control. These reports are not meant or intended for any other purpose but to uphold H₂ Hubb LLC’s business practices and to validate the reasons for our recommendations.



Approved by:



CEO, H₂HUBB LLC

