



Overland Park KS
Tywon@H₂HUBB.com
www.H₂HUBB.com

Date: 5/22/2025

H₂HUBB Official Test Report

Evaluation Introduction

This report provides a comprehensive analysis of the Nord Inhale 2.0 Hydrogen Inhalation Device from Nord Hydrogen, a company based in Sweden. H₂HUBB classifies this device as a Household-Grade, mid-to-high flow hydrogen inhalation system. It is equipped with an hydrogen electrolytic cell utilizing a PEM/SPE membrane, ensuring the production of pure hydrogen gas with only deionized or distilled water required for operation. The device supports continuous hydrogen gas production, allowing users to extend inhalation sessions up to 6 hours. We thoroughly evaluated the system's hydrogen gas output in mL/min to verify its performance. Additionally, we assessed its safety features and operations to confirm the presence of appropriate mechanisms for safe and reliable usage. Our investigation determines whether the Nord Inhale 2.0 device meets our H₂ product performance standards required for approval and recommendation by H₂HUBB. For more information about our performance standards for hydrogen inhalation systems, please visit [H₂HUBB](https://www.h2hubb.com).

H₂ Products

- Company: Nord Hydrogen
- Product Name: Nord Inhale 2.0
- Type: Pure H₂ Inhalation Device (99.9%/3N)
 - PEM/SPE
 - Mid-to- high Flow Rate
- Mfg rated H₂ Output: 600 mL/min
- URL Link: <https://nordhydrogen.com/products/nord-inhale-2-0>

Method and Procedure

- Distilled Water (used for testing): 6.0 pH
- Water Temperature: 65~70F/ 18~21C
- Reservoir Vol Size: 1.0 L/1000 mL (0.26 gals)
- H₂ Output: 600 mL/min or 49.46 mg/min (@ SATP)
- Test Location: 277 meters (909 ft elevation)
- H₂ Flow Test: mL/min, normal timing for a breathing session (1 hr)
- Test methodology:
 - [Alicat H₂ Mass Flow Meter](#)
 - [Gas Displacement](#)
- All measurements converted to SATP where applicable

Test Results

To conduct the hydrogen gas flow rate test, the system was properly assembled and filled with distilled water to the recommended level. The device was set to the 900 mL/min output mode (comprising 600 mL/min of H₂ and 300 mL/min of O₂) and activated for a 1-hour session, with a 10-minute warm-up period prior to measurement. The generated hydrogen gas was directed through a dryer, then passed through a humidity and temperature sensor before entering the Alicat Mass Flow Meter for precise measurement of molecular hydrogen output. To ensure measurement accuracy, the system was allowed to stabilize for 5–10 minutes, preventing interference from ambient air. Minor adjustments were applied to account for any flow reduction caused by moisture or resistance due to the dryer. A minimum of three tests were conducted, and the final flow rate values presented in this report represent the average of those tests.

H₂ Flow Rate Test Results at SATP:

- Device H₂ mL/min (mg/min) avg: \cong 661 mL/min (11.0 mL/sec) (54.46 mg/min)
- Device O₂ mL/min (mg/min) avg: \cong 330.51 mL/min
- Device Total H₂/O₂ mL/min avg: \cong 991.50 mL/min
- Device H₂ mL/sec per-breath Pulse Mode (effective range): \cong 44–60.5 mL/sec
 - (Combined standard and stored H₂ during enhanced inhalation)
- Pulse Mode delivery frequency: \cong 70–80%

Claimed Mfg'r's H₂ mL/min (mg/min) confirmed: Yes

H₂HUBB Hydrogen Flow Rate Assessment

- H₂HUBB's hydrogen gas flow rate test results confirm the manufacturer's claims regarding the Nord Inhale 2.0's gas production capabilities. Specifically, at standard atmospheric pressure (SATP), the device produces 661 mL/min of hydrogen gas, aligning with the manufacturer's stated output. These findings not only validate the device's specifications under real-use conditions but also reflect consistent performance. Additionally, our analysis of the Pulse Mode feature shows that it enhances per-breath hydrogen delivery to 44–60.5 mL/sec in 70–80% of breaths, effectively doubling—and in some instances tripling—the inhaled hydrogen concentration. When used with a nasal cannula, this corresponds to an average inhaled hydrogen concentration of approximately 4.0–8.8% per enhanced inhalation. These results exceed H₂HUBB's minimum performance standards and qualify the Nord Inhale 2.0 as a Level 4 hydrogen inhalation device within our performance ranking system.

INTERNAL BREAKDOWN AND PERFORMANCE:

Manufacturer's Rated Electrical Values:

- Type of device/electrolytic cell
 - Pure H₂: PEM/SPE membrane
- Applied volts:
 - 12 volts
- Total Amps:
 - 27 amps
- Total watts:
 - 324 watts

Confirmed Electrical Values

- Cell Configuration:
 - Four cells in series
- Electrolysis volts at cell:
 - 6.72 volts (DC)
 - 1.68 volts (DC) per cell
- Electrolysis amps at cell:
 - 23.46 amps (DC) per cell
 - 93.84 effective amps (DC)
- Electrolysis watts at cell:
 - 157.65 watts

H₂ Production: (Based on measured amperage @SATP)

- Total Theoretical Max H₂ production (@ 100% cell efficiency)
 - Total: 714.21 mL/min (58.88 mg/min)
- Measured H₂ production
 - 661 mL/min (54.46 mg/min)
- Electrolytic cell efficiency
 - 93%

Product Assessment

Functionality:

- **Power input/Power cord:**
 - Located on the back of the system; supplies power to the device.
- **LED Digital Display and Control Panel**
 - Displays the session duration
 - Shows the combined hydrogen and oxygen gas production in mL/min
- **Output Flow:**
 - Allows the user to select hydrogen/oxygen gas production settings: 150, 300, 450, 600, 750, or 900 mL/min
- **Power Start/Stop Button:**
 - Initiates electrolysis for hydrogen inhalation
 - Pressing the button while the device is running will stop gas production
- **Timer:**
 - Adjusts session time in 1-hour increments, up to 6 hours
- **Pulse Mode:**
 - Delivers a small additional quantity of hydrogen gas during each inhalation
- **Sleep Mode:**
 - Dims the digital display to reduce ambient light during use
- **Reservoir (1.0 L or 0.26 gal):**
 - Requires 1.0 liter of distilled water
- **H₂ Port (1x):**
 - Outputs hydrogen gas for single-user inhalation
- **O₂ Vent (1x):**
 - Vents oxygen gas produced during electrolysis
- **Pulse Port (1x):**
 - Releases stored hydrogen gas to support enhanced inhalation during Pulse Mode
- **Drain Port:**
 - Allows the user to empty the distilled water reservoir using a special drain fitting

Product Safety

Safety Components:

- The system has 5 fundamental safety mechanisms for ensuring the device's safety.
 - Low-water protection
 - Protects cells from excessive heat
 - Large distilled water reservoir
 - Protects cells from excessive heat
 - Internal Fans
 - Prevents hydrogen gas build-up in case of leaks and may also aid in preventing overheating
 - Internal gas separator
 - The apparatus helps to improve H₂ gas purity.
 - Internal deionization resin filters
 - Improves gas purity and reduces ions (mineral, metal, etc.)
 - Heat Vents
 - Prevents excessive heat in the system

The system theoretically should only be combustible at the tip of the nasal cannula as the system produces >99% pure hydrogen gas. As with all inhalation devices that produce pure hydrogen gas, care should be taken to avoid exposing the cannula tip to any source of ignition (such as an open flame or a spark) which could result in the combustion of the gas.

Overall Opinion

The Nord Inhale 2.0 Hydrogen Inhalation Device has been confirmed through our testing as a well-engineered system for hydrogen inhalation. The manufacturer rates the device to deliver 600 mL/min of 99.9% pure hydrogen gas under standard operating conditions. Our testing verified these claims, demonstrating that the device consistently achieves the specified hydrogen output, with measured results aligning with the rated flow rate. Additionally, our evaluation confirmed the functionality of the device's Pulse Mode, which enhances per-breath hydrogen delivery and effectively increases inhaled hydrogen concentration in alignment with the manufacturer's intended design.

Hydrogen gas output flow rates are a critical performance parameter for inhalation devices. At H₂HUBB, the minimum standard for hydrogen generators or inhalation units (whether pure hydrogen, oxyhydrogen, or H₂ mixed with air) is 120 mL/min of H₂. This rate corresponds to approximately 0.7-1.3% H₂ at typical resting breathing rates (4-6 L/min) when using a nasal cannula for an average adult. Scientific studies on molecular hydrogen inhalation therapy generally utilize concentrations between 0.5% and 4% or more at resting breathing rates, a range that has been shown to provide therapeutic benefits. Given these findings, H₂HUBB establishes 120 mL/min of H₂ as the baseline requirement for hydrogen inhalation devices to ensure effectiveness. The Nord Inhale 2.0 H₂ inhalation device significantly exceeds this minimum standard, delivering performance well within the therapeutic range.

The Nord Inhale 2.0 is equipped with an advanced PEM/SPE electrolytic cell system configured in a quad-cell arrangement, engineered to consistently produce high-purity hydrogen gas (>99.9%). During our testing, the system's electrical values were recorded at 6.72V and 23.46A at full production capacity, indicating that each cell operates at 1.68V drawing 23.46 amps. Based on these values, our measured hydrogen output of approximately 661 mL/min (54.46 mg/min) not only confirms but exceeds the manufacturer's stated output of 600 mL/min, aligning well with our findings. Theoretical calculations based on the measured electrical input suggest a maximum hydrogen gas output of approximately 714.21 mL/min at SATP, assuming 100% cell efficiency. This indicates the system operates at an impressive 93% efficiency, which is exceptional for a PEM/SPE-based hydrogen inhalation device. Our verified output of 661 mL/min under SATP conditions represents a 10.17% increase over the manufacturer's claim, demonstrating that the device performs slightly above expectations. This suggests the manufacturer has taken a conservative approach in its specifications, opting to under-promise and over-deliver—an approach that deserves commendation. Based on current scientific research and resting breathing rates, we conclude that the Nord Inhale 2.0's hydrogen flow rate is more than adequate to deliver therapeutic levels of hydrogen gas, achieving an inhaled H₂ concentration of approximately 4.0–4.4% for the average adult using a nasal cannula. **According to our flow-rate test results, the product will be featured on our website as a Level 4 hydrogen inhalation device.** You can view the meaning of this ranking [here](#).

The device features a function known as Pulse Mode, designed to introduce a small quantity of additional hydrogen gas into the nasal cavity with each inhalation. To activate this function, the user must press the Pulse Mode button. Once activated, the system attempts to synchronize with the user's inhalation pattern and delivers approximately 11–22 mL of hydrogen gas per breath, depending on the duration of the inhalation (calculated as: $661 \text{ mL/min} \div 60 \text{ sec} \times 1\text{--}2 \text{ sec} = 11.0\text{--}22.0 \text{ mL}$). It is claimed that the Pulse Mode function can double the inhaled hydrogen for individuals using the device during inhalation. Based on our confirmed flow rate measurement of approximately 661 mL/min (or 54.47 mg/min) at SATP, the system provides a continuous hydrogen delivery of 11.0 mL/sec, corresponding to an inhaled concentration of approximately 4.0–4.4% for an average adult male at rest. Theoretically, if Pulse Mode enables a momentary doubling of the hydrogen volume per inhalation—raising the delivery to 22–44 mL/sec—the system could achieve a peak inhalation concentration of up to 8.8%. This would equate to an inhaled hydrogen flow rate of up to 1322 mL/min, although this is only momentarily achievable.

The system includes a H₂ gas reservoir that stores hydrogen during exhalation and releases it when either:

- the reservoir pressure reaches 5.68 psi (0.4 kg/cm²), or
- a negative pressure (indicative of inhalation) is detected at the nasal cannula tip.

Based on discussions with the manufacturer and our own measurements, the device continues to deliver the standard 661 mL/min of hydrogen continuously, and releases the stored gas only under the conditions noted above. We conducted timing tests and observed that the reservoir typically releases its contents every 4.5 to 5 seconds.

Using the measured flow rate:

- $661 \text{ mL/min} \div 60 \text{ sec} = 11.0 \text{ mL/sec}$
- $11.0 \text{ mL/sec} \times 4.5\text{--}5 \text{ sec} = 49.5\text{--}55.0 \text{ mL}$

This aligns well with our gas displacement measurements, which showed a released volume of ≥ 40 mL per cycle.

To understand the practical impact, it's important to consider standard respiratory physiology:

- Tidal volume: $\sim 250\text{--}500$ mL
- Ventilation rate: $\sim 4\text{--}6$ L/min
- Inhalation/exhalation ratio: 1:2 (about 1–2 sec inhalation, 2–3 sec exhalation)
- Breathing rate: $\sim 12\text{--}16$ breaths/min (1 breath every 3–5 sec)

Given this, most inhalations occur every 3–3.5 seconds. At that interval, the device stores:

- $11.0 \text{ mL/sec} \times 3\text{--}3.5 \text{ sec} = 33\text{--}38.5$ mL of additional hydrogen

When this stored hydrogen is added to the standard 11–22 mL delivered during inhalation, the total per-breath delivery increases to:

- 44–60.5 mL of hydrogen

Our tests found that the device does not release stored hydrogen with every inhalation, but does so approximately 70–80% of the time. For example, over 10 breaths:

- 7–8 inhalations include an additional 33–38.5 mL of hydrogen
- 2–3 inhalations receive only the standard 11.0 mL

Using 7 breaths at 44 mL (standard + stored) and 3 breaths at 11 mL:

- $(7 \times 44) + (3 \times 11) = 308 + 33 = 341$ mL of hydrogen delivered over 10 breaths

Assuming an average of 13 breaths per minute:

- Total H_2 per minute = $(341 \div 10) \times 13 = 443.3$ mL/min
- Equivalent mL/sec = $443.3 \div 60 = 7.39$ mL/sec
- Increase over baseline ($11 \text{ mL/sec} \times 1/3$ inhalation time = 3.67 mL/sec) = $7.39 \div 3.67 \approx 2.0\times$

Alternatively, using a maximum case (8 breaths with stored gas at 60.5 mL and 2 without):

- $(8 \times 60.5) + (2 \times 11) = 484 + 22 = 506$ mL over 10 breaths
- Total H_2 per min = $(506 \div 10) \times 13 = 657.8$ mL/min
- $657.8 \div 60 = 10.96$ mL/sec
- $10.96 \div 3.67 \approx 3\times$ increase over baseline

Thus, in ideal conditions, the system may effectively triple the inhaled hydrogen flow rate at times.

While Pulse Mode appears capable of significantly enhancing hydrogen inhalation volumes, its performance is dependent on several variables:

- User breathing patterns
- Pressure sensor sensitivity
- Reservoir refill timing

Despite variability in delivery and the fact that stored hydrogen is not released with every breath, the Pulse Mode feature demonstrates a promising advancement in hydrogen inhalation technology. Based on our testing, measured performance, and the supporting physiological variables, Pulse Mode appears capable of doubling—and in some instances tripling—the inhaled hydrogen concentration per inhalation.

While the feature could still benefit from refinement to improve timing precision and consistency, H₂HUBB is convinced of the efficacy of Pulse Mode in its current implementation to at least double the inhaled H₂ concentration. This feature represents a significant innovation that sets the Nord Inhale 2.0 apart from other hydrogen inhalation systems currently available on the global H₂ market..

The manufacturer's claims regarding the molecular hydrogen output flow rate have been validated by our tests, and the device's performance aligns well with the product's marketing materials. No safety concerns were identified, as the system appears to incorporate adequate safety measures. Overall, we are satisfied with the device's performance. The Nord Inhale 2.0 device exceeded H₂HUBB's minimum performance standards, and in our assessment, it is both safe and effective for in-home hydrogen inhalation. Based on these findings, we are confident in recommending this product to the public.

H₂ Hubb LLC disclaimer: All tests conducted and test results produced by H₂ Hubb LLC have been done according to industry-accepted practices and standards. Nevertheless, these results may not necessarily reflect test results performed by manufacturers, suppliers or third-party labs. Our test results are independent of all other parties, and testing by other parties may produce different results. We understand that many variables are involved in testing, some of which are extremely difficult to control. These reports are not meant or intended for any other purpose but to uphold H₂ Hubb LLC's business practices and to validate the reasons for our recommendations.



Approved by:

CEO, H₂HUBB LLC

